
Midmark M11 - v1.0.5
Total Cycles: 1940

Name: RITTER2
Sterilizer ID: V2313228
01/17/2023 03:13 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 191.0 F 0.0
2:00 182.7 F 3.7
4:00 195.8 F 8.1
6:00 212.3 F 13.1
8:00 227.5 F 8.3
10:00 241.0 F 13.1
12:00 240.2 F 11.7
14:00 239.9 F 10.9
16:00 253.2 F 17.7
18:00 264.6 F 24.5
20:00 270.9 F 29.5

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 30.0
1:00 272.6 F 30.1
2:00 273.0 F 30.5
3:00 271.9 F 30.2
4:00 272.5 F 30.2

Min 271.9 F 29.9
Max 273.3 F 30.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.9 F 0.0
5:00 238.2 F 0.0
10:00 227.9 F 0.0
15:00 230.7 F 0.0
20:00 230.3 F 0.0
25:00 228.1 F 0.0
30:00 229.8 F 0.0
Drying Complete

FILLING: 1:56
HEATING: 20:38
STERILIZING: 4:00
VENTING: 3:27
DRYING: 30:00
TOTAL CYCLE: 01:00:01
CYCLE COMPLETE

01/17/2023 04:12 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____