
Midmark M11 - v1.0.5
Total Cycles: 4474

Name: RITTER2
Sterilizer ID: V2313228
03/31/2025 01:56 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 116.4 F 0.0
2:00 117.3 F 0.9
4:00 117.5 F 3.1
6:00 140.9 F 6.5
8:00 174.0 F 10.6
10:00 201.5 F 4.6
12:00 215.3 F 6.5
14:00 226.4 F 10.5
16:00 228.4 F 8.2
18:00 237.9 F 11.6
20:00 236.0 F 9.5
22:00 248.7 F 15.6
24:00 259.8 F 21.7
26:00 269.8 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.5
1:00 272.5 F 30.1
2:00 271.4 F 29.4
3:00 272.3 F 30.3
4:00 271.8 F 29.6

Min 271.3 F 29.4
Max 272.6 F 30.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.8 F 0.0
5:00 213.5 F 0.0
10:00 198.5 F 0.0
15:00 192.8 F 0.0
20:00 193.2 F 0.0
25:00 184.7 F 0.0
30:00 184.2 F 0.0
Drying Complete

FILLING: 0:58
HEATING: 26:57
STERILIZING: 4:00
VENTING: 3:39
DRYING: 30:00
TOTAL CYCLE: 01:05:34
CYCLE COMPLETE

03/31/2025 03:01 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____