
Midmark M11 - v1.0.0
Total Cycles: 5157

Name: RITTER 1
Sterilizer ID: V2178198
12/04/2024 01:41 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 158.5 F 0.0
2:00 161.5 F 0.1
4:00 184.2 F 0.7
6:00 207.1 F 1.5
8:00 217.7 F 5.8
10:00 230.7 F 10.3
12:00 232.5 F 8.1
14:00 243.4 F 12.8
16:00 242.0 F 11.6
18:00 241.1 F 10.8
20:00 254.8 F 17.8
22:00 266.4 F 24.7

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.1
1:00 271.5 F 28.3
2:00 272.4 F 29.0
3:00 272.5 F 28.9
4:00 272.2 F 28.6

Min 271.3 F 28.0
Max 272.6 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.5 F 0.0
5:00 227.4 F 0.0
10:00 214.3 F 0.0
15:00 203.9 F 0.0
20:00 200.3 F 0.0
25:00 198.2 F 0.0
30:00 198.5 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 23:35
STERILIZING: 4:00
VENTING: 3:47
DRYING: 30:00
TOTAL CYCLE: 01:02:14
CYCLE COMPLETE

12/04/2024 02:43 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____