
Midmark M11 - v1.0.0
Total Cycles: 2152

Name: RITTER 1
Sterilizer ID: V2178198
07/19/2022 03:06 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 126.0 F 0.0
2:00 128.5 F 0.0
4:00 166.0 F 2.3
6:00 169.3 F 6.6
8:00 194.2 F 11.1
10:00 209.5 F 3.7
12:00 225.5 F 8.3
14:00 237.2 F 12.5
16:00 233.9 F 8.8
18:00 240.6 F 10.9
20:00 244.3 F 12.6
22:00 256.6 F 19.1
24:00 267.6 F 25.9

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.6
1:00 272.6 F 29.3
2:00 271.6 F 28.5
3:00 272.4 F 29.4
4:00 272.2 F 28.9

Min 271.4 F 28.4
Max 272.7 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.2 F 0.0
5:00 202.6 F 0.0
10:00 209.7 F 0.0
15:00 217.7 F 0.0
20:00 222.7 F 0.0
25:00 223.7 F 0.0
30:00 227.3 F 0.0
Drying Complete

FILLING: 1:01
HEATING: 25:25
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 01:03:36
CYCLE COMPLETE

07/19/2022 04:09 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____