
Midmark M11 - v1.0.0
Total Cycles: 4641

Name: RITTER 1
Sterilizer ID: V2178198
06/14/2024 08:03 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.9 F 0.0
2:00 75.7 F 0.0
4:00 83.3 F 0.1
6:00 114.4 F 0.3
8:00 149.5 F 1.6
10:00 184.5 F 4.7
12:00 192.9 F 8.3
14:00 208.5 F 11.8
16:00 213.0 F 4.2
18:00 224.8 F 8.1
20:00 235.3 F 12.4
22:00 234.0 F 8.6
24:00 242.8 F 11.7
26:00 243.9 F 12.2
28:00 255.6 F 18.4
30:00 266.2 F 24.8

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.3
1:00 272.5 F 28.8
2:00 271.7 F 28.6
3:00 272.3 F 28.8
4:00 271.7 F 28.6

Min 271.5 F 28.2
Max 272.6 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 193.6 F 0.0
5:00 204.3 F 0.0
10:00 200.6 F 0.0
15:00 213.8 F 0.0
20:00 218.7 F 0.0
25:00 229.8 F 0.0
30:00 230.8 F 0.0
Drying Complete

FILLING: 0:46
HEATING: 31:41
STERILIZING: 4:00
VENTING: 3:24
DRYING: 30:00
TOTAL CYCLE: 01:09:52
CYCLE COMPLETE

06/14/2024 09:12 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____