
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
2.0uS / 1.2ppm
CYCLE NUMBER 003687
8:30 26/05/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:15
PRESSURIZING 2:23
STERILIZING 3:18
136.0 C 220kPa 3:18
136.4 C 221kPa 3:48
136.4 C 221kPa 4:18
136.4 C 223kPa 4:48
136.4 C 222kPa 5:18
136.0 C 221kPa 5:48
136.4 C 223kPa 6:18
136.4 C 220kPa 6:48
136.4 C 220kPa 6:48
VENTING 6:49
AIR DRYING 8:18
CYCLE COMPLETE 8:21

Digital Signature #
7ED3B13AB5046EB4
