

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 004520  
11:45 13/10/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:51  
PRESSURIZING 2:55  
STERILIZING 3:52  
136.0 C 220kPa 3:52  
136.3 C 223kPa 4:23  
136.4 C 221kPa 4:53  
136.3 C 223kPa 5:23  
136.3 C 223kPa 5:53  
136.2 C 221kPa 6:23  
136.3 C 223kPa 6:53  
136.4 C 219kPa 7:23  
136.4 C 219kPa 7:23  
VENTING 7:23  
AIR DRYING 8:12  
CYCLE COMPLETE 8:17

-----  
Digital Signature #  
7835BAE73F82AC49  
-----