

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4407

Name: RITTER2  
Sterilizer ID: V2313228  
03/06/2025 02:05 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 127.1 F 0.0  
2:00 129.0 F 1.3  
4:00 140.0 F 3.8  
6:00 172.5 F 7.9  
8:00 195.6 F 12.5  
10:00 210.9 F 6.3  
12:00 224.6 F 11.0  
14:00 227.6 F 7.1  
16:00 239.7 F 13.0  
18:00 240.2 F 11.7  
20:00 253.0 F 18.2  
22:00 264.6 F 25.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 30.0  
1:00 273.0 F 30.8  
2:00 272.1 F 30.1  
3:00 271.7 F 30.3  
4:00 273.0 F 30.8

Min 271.5 F 29.8  
Max 273.1 F 30.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 201.9 F 0.0  
5:00 200.1 F 0.0  
10:00 208.3 F 0.0  
15:00 213.6 F 0.0  
20:00 213.6 F 0.0  
25:00 214.9 F 0.0  
30:00 214.5 F 0.0  
Drying Complete

FILLING: 0:58  
HEATING: 23:57  
STERILIZING: 4:00  
VENTING: 3:41  
DRYING: 30:00  
TOTAL CYCLE: 01:02:36  
CYCLE COMPLETE

03/06/2025 03:07 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_