
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 002060
7:34 20/09/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:37
PRESSURIZING 2:44
STERILIZING 3:42
136.0 C 220kPa 3:42
136.3 C 222kPa 4:12
136.3 C 223kPa 4:42
136.1 C 223kPa 5:12
136.3 C 224kPa 5:42
136.4 C 224kPa 6:12
136.3 C 223kPa 6:42
136.4 C 224kPa 7:12
136.4 C 224kPa 7:12
VENTING 7:12
AIR DRYING 8:03
CYCLE COMPLETE 38:03

Digital Signature #
3D99B3198179B0AB
