
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
2.7uS / 1.7ppm
CYCLE NUMBER 003267
10:20 14/03/2022
WRAPPED
135 C for 10 min.
CYCLE START 0:00
CONDITIONING 1:30
PRESSURIZING 3:42
STERILIZING 4:41
136.4 C 221kPa 4:41
136.4 C 224kPa 5:11
136.1 C 221kPa 5:41
136.3 C 222kPa 6:11
136.4 C 220kPa 6:41
136.2 C 221kPa 7:11
136.3 C 221kPa 7:41
136.4 C 219kPa 8:11
136.4 C 220kPa 8:41
136.4 C 222kPa 9:11
136.2 C 222kPa 9:41
136.4 C 220kPa 10:11
136.4 C 221kPa 10:41
136.4 C 221kPa 11:11
136.4 C 223kPa 11:41
136.3 C 222kPa 12:11
136.3 C 223kPa 12:41
136.2 C 220kPa 13:11
136.1 C 218kPa 13:41
136.4 C 222kPa 14:11
136.2 C 217kPa 14:41
136.2 C 217kPa 14:41
VENTING 14:41
AIR DRYING 15:29
CYCLE COMPLETE 37:14

Digital Signature #

42D182C75713C018
