
Midmark M11 - v1.0.0
Total Cycles: 4059

Name: RITTER 1
Sterilizer ID: V2178198
11/27/2023 12:04 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 187.7 F 0.0
2:00 184.6 F 2.9
4:00 198.7 F 7.8
6:00 215.3 F 13.3
8:00 227.2 F 8.1
10:00 239.5 F 13.3
12:00 239.9 F 10.7
14:00 236.1 F 8.9
16:00 250.5 F 15.6
18:00 262.1 F 22.1
20:00 272.5 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 28.4
1:00 272.1 F 28.8
2:00 271.5 F 28.8
3:00 272.1 F 28.7
4:00 272.4 F 29.0

Min 271.0 F 28.2
Max 273.0 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 189.4 F 0.0
5:00 200.1 F 0.0
10:00 207.4 F 0.0
15:00 199.6 F 0.0
20:00 194.9 F 0.0
25:00 191.4 F 0.0
30:00 188.5 F 0.0
Drying Complete

FILLING: 1:44
HEATING: 20:24
STERILIZING: 4:00
VENTING: 3:32
DRYING: 30:00
TOTAL CYCLE: 00:59:39
CYCLE COMPLETE

11/27/2023 01:03 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____