
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001543
11:51 24/10/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:08
PRESSURIZING 3:10
STERILIZING 4:12
136.4 C 225kPa 4:12
136.3 C 226kPa 4:42
136.0 C 222kPa 5:12
136.4 C 226kPa 5:42
136.4 C 227kPa 6:12
136.4 C 224kPa 6:42
136.3 C 224kPa 7:12
136.4 C 224kPa 7:42
136.4 C 224kPa 7:42
VENTING 7:42
AIR DRYING 8:41
CYCLE COMPLETE 9:55

Digital Signature #
B2C69837C682391C
