
Midmark M11 - v1.0.5
Total Cycles: 1490

Name: RITTER2
Sterilizer ID: V2313228
09/14/2022 12:07 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 130.2 F 0.0
2:00 132.1 F 1.3
4:00 153.3 F 4.2
6:00 177.0 F 8.2
8:00 205.5 F 13.3
10:00 216.7 F 7.1
12:00 229.2 F 11.9
14:00 231.6 F 8.5
16:00 240.7 F 11.5
18:00 243.7 F 12.8
20:00 255.7 F 19.2
22:00 267.0 F 26.2

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.4
1:00 272.5 F 30.2
2:00 272.1 F 29.6
3:00 271.2 F 29.5
4:00 272.4 F 30.3

Min 271.1 F 29.2
Max 272.7 F 30.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.9 F 0.0
5:00 210.1 F 0.0
10:00 232.6 F 0.0
15:00 226.7 F 0.0
20:00 219.4 F 0.0
25:00 217.6 F 0.0
30:00 215.8 F 0.0
Drying Complete

FILLING: 1:05
HEATING: 23:33
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 01:01:47
CYCLE COMPLETE

09/14/2022 01:08 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____