
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 008170
9:28 16/12/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:58
PRESSURIZING 2:59
STERILIZING 3:56
136.0 C 220kPa 3:56
136.1 C 221kPa 4:26
136.4 C 221kPa 4:56
136.4 C 224kPa 5:26
136.4 C 222kPa 5:56
136.3 C 220kPa 6:26
136.2 C 217kPa 6:56
136.3 C 222kPa 7:26
136.3 C 222kPa 7:26
VENTING 7:26
AIR DRYING 8:18
CYCLE COMPLETE 10:19

Digital Signature #
A669AAE2E8F665B6
