
Midmark M11 - v1.0.5
Total Cycles: 1856

Name: RITTER2
Sterilizer ID: V2313228
12/19/2022 01:45 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 203.8 F 0.0
2:00 197.7 F 4.3
4:00 203.3 F 10.0
6:00 220.8 F 6.2
8:00 234.3 F 12.0
10:00 236.9 F 10.3
12:00 236.8 F 9.4
14:00 251.7 F 16.8
16:00 264.2 F 24.1

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.2
1:00 272.6 F 29.6
2:00 271.6 F 29.3
3:00 272.4 F 29.6
4:00 271.6 F 29.1

Min 271.5 F 29.0
Max 272.7 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.4 F 0.0
5:00 218.5 F 0.0
10:00 248.8 F 0.0
15:00 229.5 F 0.0
20:00 232.4 F 0.0
25:00 231.1 F 0.0
30:00 227.0 F 0.0
Drying Complete

FILLING: 2:49
HEATING: 17:55
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:57:55
CYCLE COMPLETE

12/19/2022 02:42 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____