

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R750  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 000795  
13:08 11/11/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:34  
PRESSURIZING 3:36  
STERILIZING 4:38  
136.4 C 224kPa 4:38  
136.4 C 225kPa 5:08  
136.1 C 222kPa 5:38  
136.4 C 219kPa 6:08  
136.4 C 220kPa 6:38  
136.4 C 225kPa 7:08  
136.3 C 223kPa 7:38  
136.4 C 224kPa 8:08  
136.4 C 224kPa 8:08  
VENTING 8:08  
AIR DRYING 9:04  
CYCLE COMPLETE 9:12

-----  
Digital Signature #  
C3CC43101CEB966B  
-----