
Midmark M11 - v1.0.5
Total Cycles: 1391

Name: RITTER2
Sterilizer ID: V2313228
08/15/2022 11:33 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 194.7 F 0.0
2:00 182.4 F 0.8
4:00 199.6 F 5.9
6:00 217.9 F 11.0
8:00 226.1 F 7.2
10:00 239.1 F 12.9
12:00 239.5 F 11.3
14:00 239.8 F 10.6
16:00 253.2 F 17.5
18:00 264.9 F 24.4

STERILIZING
mm:ss Degrees PSI
0:00 272.1 F 28.9
1:00 272.4 F 29.7
2:00 272.1 F 29.0
3:00 272.1 F 29.5
4:00 272.5 F 29.4

Min 271.4 F 28.8
Max 273.0 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.2 F 0.0
5:00 229.2 F 0.0
10:00 220.9 F 0.0
15:00 211.7 F 0.0
20:00 213.6 F 0.0
25:00 209.8 F 0.0
30:00 208.5 F 0.0
Drying Complete

FILLING: 1:39
HEATING: 19:44
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:58:32
CYCLE COMPLETE

08/15/2022 12:31 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____