
Midmark M11 - v1.0.0
Total Cycles: 5126

Name: RITTER 1
Sterilizer ID: V2178198
11/22/2024 11:00 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 180.3 F 0.0
2:00 174.7 F 0.0
4:00 174.1 F 0.3
6:00 188.9 F 0.7
8:00 206.6 F 3.7
10:00 221.6 F 6.9
12:00 229.5 F 10.7
14:00 234.0 F 8.5
16:00 239.1 F 11.3
18:00 235.0 F 8.7
20:00 241.7 F 11.3
22:00 245.4 F 13.0
24:00 256.1 F 18.6
26:00 265.5 F 24.2
28:00 272.1 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.3
1:00 271.7 F 28.6
2:00 272.2 F 29.0
3:00 272.4 F 29.1
4:00 272.3 F 28.9

Min 271.3 F 28.2
Max 272.5 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.1 F 0.0
5:00 203.7 F 0.0
10:00 207.0 F 0.0
15:00 225.2 F 0.0
20:00 234.0 F 0.0
25:00 233.4 F 0.0
30:00 229.0 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 28:14
STERILIZING: 4:00
VENTING: 4:02
DRYING: 30:00
TOTAL CYCLE: 01:07:05
CYCLE COMPLETE

11/22/2024 12:07 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____