
Midmark M11 - v1.0.5
Total Cycles: 2554

Name: RITTER2
Sterilizer ID: V2313228
06/22/2023 01:47 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 118.4 F 0.0
2:00 121.1 F 0.0
4:00 138.9 F 2.3
6:00 169.9 F 6.0
8:00 194.0 F 10.1
10:00 213.8 F 7.2
12:00 224.1 F 8.3
14:00 235.0 F 13.1
16:00 235.8 F 10.0
18:00 236.9 F 9.1
20:00 247.0 F 14.2
22:00 258.4 F 20.4
24:00 268.8 F 27.1

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.0
1:00 272.6 F 29.8
2:00 272.0 F 29.2
3:00 271.2 F 29.0
4:00 272.3 F 29.8

Min 271.1 F 28.8
Max 272.8 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 189.5 F 0.0
5:00 215.1 F 0.0
10:00 215.7 F 0.0
15:00 219.8 F 0.0
20:00 219.0 F 0.0
25:00 217.6 F 0.0
30:00 216.7 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 25:13
STERILIZING: 4:00
VENTING: 3:20
DRYING: 30:00
TOTAL CYCLE: 01:03:29
CYCLE COMPLETE

06/22/2023 02:50 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____