
Midmark M11 - v1.0.0
Total Cycles: 4942

Name: RITTER 1
Sterilizer ID: V2178198
09/26/2024 12:26 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 108.0 F 0.0
2:00 122.5 F 0.0
4:00 129.5 F 0.1
6:00 155.4 F 0.2
8:00 178.5 F 0.8
10:00 197.7 F 4.1
12:00 216.1 F 7.8
14:00 225.5 F 11.6
16:00 225.6 F 6.4
18:00 237.4 F 11.6
20:00 235.5 F 8.8
22:00 240.6 F 10.5
24:00 246.8 F 13.5
26:00 258.2 F 19.6
28:00 267.7 F 25.7

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.3
1:00 272.3 F 28.9
2:00 272.7 F 29.1
3:00 272.3 F 28.7
4:00 271.8 F 28.4

Min 271.4 F 28.2
Max 272.8 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.1 F 0.0
5:00 189.4 F 0.0
10:00 183.3 F 0.0
15:00 187.0 F 0.0
20:00 191.5 F 0.0
25:00 196.7 F 0.0
30:00 197.3 F 0.0
Drying Complete

FILLING: 0:46
HEATING: 29:35
STERILIZING: 4:00
VENTING: 3:47
DRYING: 30:00
TOTAL CYCLE: 01:08:08
CYCLE COMPLETE

09/26/2024 01:34 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____