

-----  
Midmark M11 - v1.0.5  
Total Cycles: 573

Name: RITTER2  
Sterilizer ID: V2313228  
11/24/2021 12:18 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 194.4 F 0.0  
2:00 190.4 F 3.8  
4:00 191.3 F 8.9  
6:00 214.1 F 7.1  
8:00 226.9 F 9.9  
10:00 230.4 F 8.2  
12:00 241.5 F 12.0  
14:00 245.2 F 13.5  
16:00 258.3 F 20.6  
18:00 269.9 F 28.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.1  
1:00 272.6 F 29.8  
2:00 271.9 F 29.1  
3:00 271.9 F 29.8  
4:00 272.5 F 29.7  
  
Min 271.4 F 29.0  
Max 272.7 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 207.0 F 0.0  
5:00 224.3 F 0.0  
10:00 238.1 F 0.0  
15:00 227.1 F 0.0  
20:00 224.8 F 0.0  
25:00 223.5 F 0.0  
30:00 225.0 F 0.0  
  
Drying Complete

FILLING: 1:43  
HEATING: 18:55  
STERILIZING: 4:00  
VENTING: 3:09  
DRYING: 30:00  
TOTAL CYCLE: 00:57:47  
CYCLE COMPLETE

11/24/2021 01:15 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_