
Midmark M11 - v1.0.0
Total Cycles: 863

Name: RITTER 1
Sterilizer ID: V2178198
06/07/2021 08:07 AM

BEGIN PACKS CYCLE
Temp: 250 Degrees F
Time: 30 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 176.6 F 0.0
2:00 167.1 F 1.3
4:00 178.3 F 4.8
6:00 213.3 F 2.2
8:00 217.2 F 5.9
10:00 222.9 F 4.2
12:00 227.8 F 5.0
14:00 236.8 F 9.1
16:00 249.0 F 14.8

STERILIZING
mm:ss Degrees PSI
0:00 251.2 F 16.2
1:00 252.4 F 16.5
2:00 251.7 F 16.5
3:00 252.2 F 16.4
4:00 251.4 F 16.5
5:00 252.7 F 16.5
6:00 251.4 F 16.1
7:00 252.7 F 16.8
8:00 252.6 F 16.5
9:00 251.4 F 16.2
10:00 252.1 F 16.7
11:00 253.0 F 16.6
12:00 252.1 F 16.4
13:00 251.3 F 16.1
14:00 252.2 F 16.8
15:00 253.0 F 16.7
16:00 252.4 F 16.5
17:00 251.5 F 16.2
18:00 251.3 F 16.4
19:00 252.9 F 16.8
20:00 252.6 F 16.6
21:00 252.2 F 16.4
22:00 251.5 F 16.2
23:00 251.4 F 16.5
24:00 252.6 F 16.8
25:00 252.7 F 16.6
26:00 252.3 F 16.4
27:00 251.5 F 16.2
28:00 251.3 F 16.4
29:00 252.5 F 16.8
30:00 253.0 F 16.7

Min 251.2 F 16.1
Max 253.2 F 16.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.8 F 0.0
5:00 209.3 F 0.0
10:00 216.0 F 0.0
15:00 202.6 F 0.0
20:00 196.5 F 0.0
Dry Cycle Aborted

FILLING: 1:33
HEATING: 17:03
STERILIZING: 30:00
VENTING: 2:59

DRYING: 24:46
TOTAL CYCLE: 01:16:21

06/07/2021 09:23 AM
ACCEPTED: X
REJECTED:
Operator: U01