
Midmark M11 - v1.0.0
Total Cycles: 1762

Name: RITTER 1
Sterilizer ID: V2178198
03/29/2022 01:38 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 127.2 F 0.0
2:00 156.1 F 0.1
4:00 182.3 F 2.4
6:00 194.8 F 5.5
8:00 221.1 F 10.0
10:00 226.2 F 6.2
12:00 229.3 F 9.3
14:00 236.6 F 9.7
16:00 241.7 F 11.8
18:00 239.6 F 10.0
20:00 253.0 F 16.7
22:00 264.4 F 23.5

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.3
1:00 272.3 F 28.7
2:00 271.9 F 28.7
3:00 272.1 F 28.6
4:00 272.0 F 28.7

Min 271.5 F 28.2
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.1 F 0.0
5:00 218.0 F 0.0
10:00 222.4 F 0.0
15:00 227.6 F 0.0
20:00 236.1 F 0.0
25:00 228.9 F 0.0
30:00 228.6 F 0.0
Drying Complete

FILLING: 0:55
HEATING: 23:56
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 01:01:56
CYCLE COMPLETE

03/29/2022 02:39 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____