
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 008086
9:50 21/11/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:59
PRESSURIZING 2:02
STERILIZING 3:02
136.0 C 220kPa 3:02
136.2 C 221kPa 3:32
136.4 C 224kPa 4:02
136.4 C 222kPa 4:32
136.4 C 221kPa 5:02
136.4 C 218kPa 5:32
136.2 C 222kPa 6:02
136.2 C 220kPa 6:32
136.2 C 220kPa 6:32
VENTING 6:32
AIR DRYING 7:21
CYCLE COMPLETE 20:50

Digital Signature #
4830B47406C8F6A7
