
Midmark M11 - v1.0.5
Total Cycles: 2241

Name: RITTER2
Sterilizer ID: V2313228
04/12/2023 12:51 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 209.7 F 0.0
2:00 201.6 F 4.5
4:00 211.2 F 10.7
6:00 225.7 F 7.3
8:00 238.8 F 13.6
10:00 241.9 F 12.3
12:00 242.6 F 12.0
14:00 256.8 F 19.4
16:00 268.9 F 27.1

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.9
1:00 272.3 F 29.4
2:00 271.3 F 29.0
3:00 272.4 F 29.6
4:00 271.6 F 28.9

Min 271.2 F 28.8
Max 272.7 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.7 F 0.0
5:00 230.9 F 0.0
10:00 243.3 F 0.0
15:00 235.4 F 0.0
20:00 227.6 F 0.0
25:00 226.3 F 0.0
30:00 224.4 F 0.0
Drying Complete

FILLING: 2:30
HEATING: 17:03
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:56:49
CYCLE COMPLETE

04/12/2023 01:47 PM
ACCEPTED:
REJECTED:

Reason: _____
Operator: _____