
Midmark M11 - v1.0.0
Total Cycles: 1151

Name: RITTER 1
Sterilizer ID: V2178198
09/08/2021 01:24 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 159.5 F 0.0
2:00 157.8 F 1.8
4:00 173.6 F 4.8
6:00 191.9 F 8.7
8:00 213.7 F 13.2
10:00 226.7 F 7.8
12:00 237.6 F 12.5
14:00 234.9 F 9.4
16:00 240.3 F 9.8
18:00 245.3 F 13.4
20:00 256.8 F 19.5
22:00 267.2 F 26.0

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.9
1:00 272.9 F 29.7
2:00 272.1 F 29.0
3:00 271.8 F 29.2
4:00 272.8 F 29.7

Min 271.6 F 28.8
Max 273.0 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 214.6 F 0.0
5:00 223.5 F 0.0
10:00 204.0 F 0.0
Dry Cycle Aborted

FILLING: 1:19
HEATING: 23:31
STERILIZING: 4:00
VENTING: 3:23
DRYING: 11:37
TOTAL CYCLE: 00:43:50

09/08/2021 02:08 PM
ACCEPTED: X
REJECTED:
Operator: U01