

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3808

Name: RITTER2  
Sterilizer ID: V2313228  
07/23/2024 11:26 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 104.2 F 0.0  
2:00 105.9 F 0.0  
4:00 121.9 F 0.4  
6:00 148.3 F 3.1  
8:00 173.9 F 6.3  
10:00 195.8 F 10.1  
12:00 207.4 F 9.1  
14:00 216.8 F 6.5  
16:00 227.4 F 10.4  
18:00 232.1 F 8.5  
20:00 239.8 F 11.8  
22:00 238.2 F 10.0  
24:00 250.2 F 15.9  
26:00 260.7 F 21.9  
28:00 270.6 F 28.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 29.0  
1:00 272.9 F 29.7  
2:00 272.0 F 29.0  
3:00 272.0 F 29.4  
4:00 272.8 F 29.7

Min 271.6 F 28.9  
Max 272.9 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 196.3 F 0.0  
5:00 191.8 F 0.0  
10:00 222.9 F 0.0  
15:00 223.5 F 0.0  
20:00 219.4 F 0.0  
25:00 221.3 F 0.0  
30:00 219.7 F 0.0  
Drying Complete

FILLING: 0:53  
HEATING: 28:52  
STERILIZING: 4:00  
VENTING: 3:41  
DRYING: 30:00  
TOTAL CYCLE: 01:07:27  
CYCLE COMPLETE

07/23/2024 12:33 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_