

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4228

Name: RITTER 1  
Sterilizer ID: V2178198  
01/24/2024 06:11 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 168.9 F 0.0  
2:00 186.1 F 4.5  
4:00 205.4 F 10.1  
6:00 224.8 F 5.3  
8:00 236.7 F 11.8  
10:00 236.8 F 9.5  
12:00 237.3 F 8.6  
14:00 250.4 F 15.3  
16:00 263.1 F 22.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.9 F 28.1  
1:00 272.3 F 28.4  
2:00 272.3 F 28.6  
3:00 271.7 F 28.6  
4:00 271.4 F 28.0  
  
Min 271.0 F 28.0  
Max 272.5 F 28.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 181.5 F 0.0  
5:00 219.3 F 0.0  
10:00 209.5 F 0.0  
15:00 216.6 F 0.0  
20:00 203.8 F 0.0  
25:00 188.6 F 0.0  
30:00 183.9 F 0.0  
Drying Complete

FILLING: 1:47  
HEATING: 17:52  
STERILIZING: 4:00  
VENTING: 3:18  
DRYING: 30:00  
TOTAL CYCLE: 00:56:57  
CYCLE COMPLETE

01/24/2024 07:07 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_