
Midmark M11 - v1.0.5
Total Cycles: 1113

Name: RITTER2
Sterilizer ID: V2313228
05/25/2022 08:45 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 152.5 F 0.0
2:00 151.1 F 1.0
4:00 165.2 F 3.9
6:00 185.5 F 8.3
8:00 211.8 F 13.4
10:00 221.7 F 8.1
12:00 234.9 F 13.2
14:00 236.7 F 10.3
16:00 235.2 F 8.3
18:00 248.7 F 15.1
20:00 260.4 F 21.8
22:00 271.2 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.2
1:00 272.6 F 29.7
2:00 271.1 F 29.1
3:00 272.7 F 29.8
4:00 271.4 F 29.1

Min 271.0 F 29.0
Max 272.8 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.7 F 0.0
5:00 225.0 F 0.0
10:00 229.5 F 0.0
15:00 227.7 F 0.0
20:00 230.9 F 0.0
25:00 232.3 F 0.0
30:00 228.2 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 22:43
STERILIZING: 4:00
VENTING: 3:03
DRYING: 30:00
TOTAL CYCLE: 01:00:40
CYCLE COMPLETE

05/25/2022 09:45 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____