
Midmark M11 - v1.0.5
Total Cycles: 1870

Name: RITTER2
Sterilizer ID: V2313228
12/21/2022 10:10 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 192.4 F 0.0
2:00 189.9 F 2.8
4:00 190.2 F 7.7
6:00 211.7 F 13.4
8:00 227.9 F 9.0
10:00 235.2 F 9.4
12:00 242.5 F 13.0
14:00 242.5 F 12.0
16:00 255.6 F 19.0
18:00 267.9 F 26.4

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.1
1:00 271.8 F 29.1
2:00 272.0 F 29.5
3:00 271.8 F 29.7
4:00 271.6 F 29.1

Min 271.3 F 29.0
Max 272.1 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.4 F 0.0
5:00 229.1 F 0.0
10:00 221.9 F 0.0
15:00 231.6 F 0.0
20:00 228.4 F 0.0
25:00 227.1 F 0.0
30:00 229.1 F 0.0
Drying Complete

FILLING: 1:53
HEATING: 19:13
STERILIZING: 4:00
VENTING: 3:04
DRYING: 30:00
TOTAL CYCLE: 00:58:10
CYCLE COMPLETE

12/21/2022 11:08 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____