
Midmark M11 - v1.0.0
Total Cycles: 4902

Name: RITTER 1
Sterilizer ID: V2178198
09/12/2024 03:46 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 147.6 F 0.0
2:00 152.8 F 0.2
4:00 180.8 F 0.7
6:00 200.8 F 3.4
8:00 216.4 F 7.5
10:00 225.4 F 11.9
12:00 228.8 F 7.2
14:00 240.9 F 12.5
16:00 239.3 F 10.3
18:00 236.3 F 8.3
20:00 250.6 F 15.3
22:00 262.0 F 21.8
24:00 272.3 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.0
1:00 272.4 F 28.8
2:00 272.0 F 28.3
3:00 271.3 F 27.9
4:00 272.1 F 28.7

Min 271.3 F 27.9
Max 272.5 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.9 F 0.0
5:00 206.1 F 0.0
10:00 202.6 F 0.0
15:00 203.5 F 0.0
20:00 208.9 F 0.0
25:00 212.4 F 0.0
30:00 211.2 F 0.0
Drying Complete

FILLING: 0:55
HEATING: 24:34
STERILIZING: 4:00
VENTING: 3:42
DRYING: 30:00
TOTAL CYCLE: 01:03:11
CYCLE COMPLETE

09/12/2024 04:49 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____