

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3251

Name: RITTER2  
Sterilizer ID: V2313228  
12/22/2023 01:13 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 146.7 F 0.0  
2:00 149.7 F 1.8  
4:00 168.6 F 5.1  
6:00 191.6 F 9.5  
8:00 213.3 F 8.3  
10:00 230.8 F 8.5  
12:00 239.4 F 13.4  
14:00 239.1 F 10.4  
16:00 235.9 F 8.5  
18:00 249.6 F 15.1  
20:00 260.8 F 21.5  
22:00 271.2 F 28.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.0 F 28.7  
1:00 272.9 F 29.5  
2:00 272.4 F 29.7  
3:00 271.6 F 29.3  
4:00 271.7 F 29.0

Min 271.0 F 28.5  
Max 273.0 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 181.4 F 0.0  
5:00 250.2 F 0.0  
10:00 208.3 F 0.0  
15:00 208.3 F 0.0  
20:00 210.4 F 0.0  
25:00 201.3 F 0.0  
30:00 186.6 F 0.0  
Drying Complete

FILLING: 1:10  
HEATING: 22:44  
STERILIZING: 4:00  
VENTING: 3:31  
DRYING: 30:00  
TOTAL CYCLE: 01:01:24  
CYCLE COMPLETE

12/22/2023 02:14 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_