

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2604

Name: RITTER 1  
Sterilizer ID: V2178198  
11/23/2022 01:12 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 193.0 F 0.0  
2:00 189.1 F 4.9  
4:00 193.3 F 9.7  
6:00 216.2 F 6.0  
8:00 225.5 F 9.4  
10:00 234.0 F 9.1  
12:00 241.9 F 12.3  
14:00 241.0 F 10.9  
16:00 254.4 F 17.7  
18:00 266.1 F 24.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.4  
1:00 272.6 F 29.2  
2:00 272.1 F 28.7  
3:00 271.5 F 28.5  
4:00 272.5 F 29.3  
  
Min 271.3 F 28.3  
Max 272.7 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 210.6 F 0.0  
5:00 226.1 F 0.0  
10:00 233.6 F 0.0  
15:00 229.2 F 0.0  
20:00 228.1 F 0.0  
25:00 229.5 F 0.0  
30:00 230.0 F 0.0  
Drying Complete

FILLING: 1:28  
HEATING: 19:40  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 00:58:19  
CYCLE COMPLETE

11/23/2022 02:10 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_