
Midmark M11 - v1.0.0
Total Cycles: 2178

Name: RITTER 1
Sterilizer ID: V2178198
07/26/2022 11:33 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 197.0 F 0.0
2:00 200.0 F 4.2
4:00 208.0 F 9.1
6:00 225.3 F 6.0
8:00 231.7 F 10.2
10:00 237.2 F 8.1
12:00 243.8 F 13.3
14:00 242.5 F 11.9
16:00 255.4 F 18.5
18:00 266.6 F 25.2

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.6
1:00 272.5 F 29.3
2:00 271.6 F 28.5
3:00 272.2 F 29.3
4:00 272.3 F 29.1

Min 271.4 F 28.4
Max 272.8 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.6 F 0.0
5:00 222.7 F 0.0
10:00 230.9 F 0.0
15:00 231.5 F 0.0
20:00 224.0 F 0.0
25:00 221.4 F 0.0
30:00 221.4 F 0.0
Drying Complete

FILLING: 2:31
HEATING: 19:37
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:59:23
CYCLE COMPLETE

07/26/2022 12:32 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____