

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2314

Name: RITTER 1  
Sterilizer ID: V2178198  
09/12/2022 04:34 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 189.0 F 0.0  
2:00 188.7 F 4.3  
4:00 201.7 F 9.0  
6:00 219.3 F 8.1  
8:00 230.4 F 9.1  
10:00 238.6 F 9.8  
12:00 241.6 F 12.2  
14:00 240.9 F 10.9  
16:00 254.1 F 17.8  
18:00 265.9 F 24.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.8  
1:00 272.6 F 29.3  
2:00 271.7 F 28.9  
3:00 272.4 F 29.3  
4:00 271.5 F 28.8  
  
Min 271.5 F 28.6  
Max 272.8 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 209.6 F 0.0  
5:00 239.2 F 0.0  
10:00 232.0 F 0.0  
15:00 230.1 F 0.0  
20:00 228.1 F 0.0  
25:00 228.7 F 0.0  
30:00 228.7 F 0.0  
Drying Complete

FILLING: 1:38  
HEATING: 19:38  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 00:58:26  
CYCLE COMPLETE

09/12/2022 05:32 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_