
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000571
11:45 09/08/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:27
PRESSURIZING 2:37
STERILIZING 3:43
136.0 C 225kPa 3:43
136.4 C 233kPa 4:13
136.2 C 230kPa 4:43
136.1 C 230kPa 5:13
136.3 C 231kPa 5:43
136.1 C 230kPa 6:13
136.2 C 229kPa 6:43
136.3 C 231kPa 7:13
136.3 C 231kPa 7:13
VENTING 7:13
AIR DRYING 8:01
CYCLE COMPLETE 23:01

Digital Signature #
F44E512FA8BC649E
