
Midmark M11 - v1.0.5
Total Cycles: 1752

Name: RITTER2
Sterilizer ID: V2313228
11/21/2022 01:56 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 181.2 F 0.0
2:00 179.9 F 0.5
4:00 198.5 F 5.0
6:00 216.9 F 10.1
8:00 227.9 F 7.6
10:00 238.6 F 13.5
12:00 241.8 F 12.3
14:00 242.0 F 11.8
16:00 256.0 F 19.1
18:00 267.9 F 26.5

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.1
1:00 272.4 F 30.3
2:00 271.8 F 29.6
3:00 272.7 F 29.9
4:00 272.4 F 30.3

Min 270.9 F 28.9
Max 273.1 F 30.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.6 F 0.0
5:00 226.7 F 0.0
10:00 237.7 F 0.0
15:00 235.7 F 0.0
20:00 231.5 F 0.0
25:00 225.9 F 0.0
30:00 226.6 F 0.0
Drying Complete

FILLING: 1:21
HEATING: 19:11
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:57:44
CYCLE COMPLETE

11/21/2022 02:53 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____