
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
0.8uS / 0.5ppm
CYCLE NUMBER 001442
14:35 09/16/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:34
PRESSURIZING 5:29
STERILIZING 6:59
136.0 C 220kPa 6:59
135.9 C 219kPa 7:29
136.4 C 223kPa 7:59
136.4 C 224kPa 8:29
136.1 C 221kPa 8:59
136.2 C 223kPa 9:29
136.5 C 223kPa 9:59
136.3 C 222kPa 10:29
VENTING 10:29
AIR DRYING 11:35
CYCLE COMPLETE 28:20

Digital Signature #
2783B19F56DCCC29
