
Midmark M11 - v1.0.5
Total Cycles: 1240

Name: RITTER2
Sterilizer ID: V2313228
06/29/2022 10:04 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 169.9 F 0.0
2:00 167.4 F 3.3
4:00 178.3 F 7.3
6:00 198.7 F 12.1
8:00 216.2 F 6.5
10:00 229.6 F 11.7
12:00 231.8 F 8.5
14:00 240.2 F 11.0
16:00 245.1 F 13.4
18:00 257.6 F 20.1
20:00 268.8 F 27.3

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.1
1:00 272.5 F 30.0
2:00 272.4 F 29.5
3:00 271.5 F 29.0
4:00 271.7 F 29.6

Min 271.3 F 28.9
Max 272.9 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.3 F 0.0
5:00 208.7 F 0.0
10:00 225.1 F 0.0
15:00 209.8 F 0.0
20:00 199.5 F 0.0
25:00 193.5 F 0.0
30:00 184.5 F 0.0
Drying Complete

FILLING: 1:06
HEATING: 21:12
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 00:59:26
CYCLE COMPLETE

06/29/2022 11:03 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____