
Midmark M11 - v1.0.0
Total Cycles: 1142

Name: RITTER 1
Sterilizer ID: V2178198
09/02/2021 02:14 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 95.2 F 0.0
2:00 99.6 F 0.0
4:00 142.9 F 0.0
6:00 153.3 F 0.2
8:00 189.5 F 0.8
10:00 200.6 F 4.5
12:00 219.6 F 7.7
14:00 230.8 F 11.4
16:00 227.5 F 6.3
18:00 239.9 F 11.6
20:00 236.4 F 8.8
22:00 241.7 F 10.6
24:00 246.2 F 13.3
26:00 258.0 F 19.6
28:00 268.6 F 26.3

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.3
1:00 272.6 F 29.0
2:00 271.7 F 28.2
3:00 272.0 F 28.8
4:00 272.4 F 28.8

Min 271.3 F 28.1
Max 272.6 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 193.4 F 0.0
5:00 202.1 F 0.0
10:00 216.9 F 0.0
15:00 223.1 F 0.0
20:00 221.7 F 0.0
25:00 222.8 F 0.0
30:00 224.0 F 0.0
Drying Complete

FILLING: 1:40
HEATING: 29:13
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 01:08:07
CYCLE COMPLETE

09/02/2021 03:22 PM
ACCEPTED: X
REJECTED:
Operator: U02