

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 001912  
11:16 31/07/2024  
WRAPPED  
135 C for 10 min.  
CYCLE START 0:00  
CONDITIONING 2:45  
PRESSURIZING 4:51  
STERILIZING 5:46  
136.4 C 224kPa 5:46  
136.2 C 223kPa 6:16  
136.4 C 225kPa 6:46  
136.3 C 224kPa 7:16  
136.3 C 224kPa 7:46  
136.1 C 222kPa 8:16  
136.2 C 223kPa 8:46  
136.4 C 224kPa 9:16  
136.4 C 222kPa 9:46  
136.2 C 224kPa 10:16  
136.2 C 222kPa 10:46  
136.4 C 225kPa 11:16  
136.2 C 224kPa 11:46  
136.4 C 225kPa 12:16  
136.4 C 224kPa 12:46  
136.5 C 225kPa 13:16  
136.4 C 224kPa 13:46  
136.4 C 221kPa 14:16  
136.4 C 225kPa 14:46  
136.4 C 221kPa 15:16  
136.4 C 221kPa 15:46  
136.4 C 221kPa 15:46  
VENTING 15:46  
AIR DRYING 16:37  
CYCLE COMPLETE 16:46  
-----

Digital Signature #  
CC834ABE18708C70  
-----