
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 006036
11:59 24/07/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:40
PRESSURIZING 3:42
STERILIZING 4:42
136.0 C 220kPa 4:42
136.4 C 222kPa 5:12
136.1 C 222kPa 5:42
136.4 C 222kPa 6:12
136.2 C 223kPa 6:42
136.4 C 220kPa 7:12
136.4 C 222kPa 7:42
136.4 C 224kPa 8:12
136.4 C 224kPa 8:12
VENTING 8:12
AIR DRYING 9:07
CYCLE COMPLETE 18:49

Digital Signature #
38E2ED5F580D3907
