
Mississauga, ON
Canada

STATIM 5000 S501R751
SN 511020J00004
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000819
13:09 28/03/2022
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:47
PRESSURIZING 5:31
STERILIZING 6:54
136.0 C 220kPa 6:54
136.4 C 222kPa 7:24
136.1 C 220kPa 7:54
136.4 C 224kPa 8:24
136.4 C 222kPa 8:54
136.4 C 222kPa 9:24
136.0 C 219kPa 9:54
136.2 C 222kPa 10:24
136.2 C 222kPa 10:24
VENTING 10:24
AIR DRYING 11:23
CYCLE COMPLETE 38:18

Digital Signature #
2873094AEE010B43
