
Midmark M11 - v1.0.5
Total Cycles: 2086

Name: RITTER2
Sterilizer ID: V2313228
02/23/2023 01:35 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 165.1 F 0.0
2:00 161.4 F 1.2
4:00 188.8 F 4.8
6:00 209.5 F 9.8
8:00 223.8 F 7.2
10:00 231.6 F 10.6
12:00 231.4 F 7.9
14:00 240.4 F 11.0
16:00 246.3 F 14.0
18:00 259.2 F 21.1
20:00 270.8 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 29.1
1:00 272.5 F 29.8
2:00 271.4 F 29.0
3:00 272.7 F 29.8
4:00 271.5 F 29.0

Min 271.2 F 28.9
Max 272.7 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 177.9 F 0.0
5:00 214.7 F 0.0
10:00 229.6 F 0.0
15:00 227.5 F 0.0
20:00 227.1 F 0.0
25:00 229.3 F 0.0
30:00 228.7 F 0.0
Drying Complete

FILLING: 0:59
HEATING: 20:36
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:58:48
CYCLE COMPLETE

02/23/2023 02:33 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____