
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 005771
14:18 15/06/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:51
PRESSURIZING 1:57
STERILIZING 2:59
136.0 C 220kPa 2:59
136.2 C 223kPa 3:29
136.4 C 222kPa 3:59
136.2 C 224kPa 4:29
136.3 C 223kPa 4:59
136.4 C 218kPa 5:29
136.4 C 223kPa 5:59
136.4 C 225kPa 6:29
136.4 C 225kPa 6:29
VENTING 6:29
AIR DRYING 7:19
CYCLE COMPLETE 7:30

Digital Signature #
BEAC0860AFB72769
