

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5265

Name: RITTER 1  
Sterilizer ID: V2178198  
01/23/2025 09:38 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 71.1 F 0.0  
2:00 93.1 F 0.0  
4:00 86.7 F 0.1  
6:00 128.1 F 0.9  
8:00 160.3 F 3.8  
10:00 182.0 F 7.1  
12:00 208.8 F 10.3  
14:00 219.1 F 7.0  
16:00 221.4 F 6.0  
18:00 232.3 F 9.7  
20:00 238.5 F 10.3  
22:00 238.3 F 10.2  
24:00 237.0 F 8.3  
26:00 247.4 F 13.8  
28:00 258.7 F 20.0  
30:00 268.9 F 26.5

STERILIZING  
mm:ss Degrees PSI  
0:00 272.0 F 28.2  
1:00 272.2 F 28.8  
2:00 272.5 F 28.5  
3:00 271.8 F 28.5  
4:00 272.8 F 28.9

Min 271.5 F 28.1  
Max 273.0 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 206.9 F 0.0  
5:00 202.3 F 0.0  
10:00 211.1 F 0.0  
15:00 199.1 F 0.0  
20:00 213.6 F 0.0  
25:00 222.4 F 0.0  
30:00 226.6 F 0.0  
Drying Complete

FILLING: 0:42  
HEATING: 31:02  
STERILIZING: 4:00  
VENTING: 3:26  
DRYING: 30:00  
TOTAL CYCLE: 01:09:11  
CYCLE COMPLETE

01/23/2025 10:47 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_