
Midmark M11 - v1.0.0
Total Cycles: 5379

Name: RITTER 1
Sterilizer ID: V2178198
02/28/2025 02:30 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 154.7 F 0.0
2:00 149.3 F 2.6
4:00 169.0 F 5.9
6:00 186.8 F 9.5
8:00 217.3 F 7.5
10:00 219.2 F 6.4
12:00 230.2 F 10.5
14:00 232.8 F 7.8
16:00 239.9 F 11.5
18:00 237.3 F 9.2
20:00 250.0 F 15.3
22:00 260.8 F 21.3
24:00 270.8 F 27.8

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.3
1:00 272.1 F 29.2
2:00 272.0 F 28.7
3:00 272.4 F 29.3
4:00 271.6 F 28.8

Min 270.8 F 28.2
Max 273.1 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 193.2 F 0.0
5:00 225.7 F 0.0
10:00 223.0 F 0.0
15:00 231.1 F 0.0
20:00 230.4 F 0.0
25:00 230.9 F 0.0
30:00 218.9 F 0.0
Drying Complete

FILLING: 1:09
HEATING: 24:40
STERILIZING: 4:00
VENTING: 3:53
DRYING: 30:00
TOTAL CYCLE: 01:03:42
CYCLE COMPLETE

02/28/2025 03:33 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____