

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2606

Name: RITTER2  
Sterilizer ID: V2313228  
07/05/2023 07:21 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 70.3 F 0.0  
2:00 72.8 F 0.0  
4:00 76.1 F 0.0  
6:00 99.5 F 0.0  
8:00 157.8 F 0.0  
10:00 198.3 F 0.6  
12:00 209.9 F 2.8  
14:00 215.1 F 6.0  
16:00 224.0 F 9.5  
18:00 235.0 F 13.3  
20:00 233.8 F 8.6  
22:00 243.3 F 13.3  
24:00 240.4 F 11.0  
26:00 237.1 F 9.4  
28:00 250.7 F 15.8  
30:00 261.5 F 22.1  
32:00 271.7 F 28.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.7  
1:00 272.7 F 29.4  
2:00 271.6 F 28.6  
3:00 272.7 F 29.5  
4:00 271.9 F 28.8  
  
Min 271.5 F 28.5  
Max 272.8 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 196.9 F 0.0  
5:00 200.7 F 0.0  
10:00 203.8 F 0.0  
15:00 213.9 F 0.0  
20:00 219.3 F 0.0  
25:00 223.0 F 0.0  
30:00 225.5 F 0.0  
  
Drying Complete

FILLING: 0:49  
HEATING: 32:38  
STERILIZING: 4:00  
VENTING: 3:09  
DRYING: 30:00  
TOTAL CYCLE: 01:10:35  
CYCLE COMPLETE

07/05/2023 08:31 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_