
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.6uS / 0.3ppm
CYCLE NUMBER 002335
13:50 29/11/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:01
PRESSURIZING 2:06
STERILIZING 3:02
136.4 C 222kPa 3:02
136.4 C 225kPa 3:32
136.1 C 223kPa 4:02
136.4 C 226kPa 4:32
136.4 C 225kPa 5:02
136.1 C 223kPa 5:32
136.4 C 226kPa 6:02
136.4 C 226kPa 6:32
136.4 C 226kPa 6:32
VENTING 6:32
AIR DRYING 7:23
CYCLE COMPLETE 16:31

Digital Signature #
0074FC259B358310
