
Midmark M11 - v1.0.5
Total Cycles: 4372

Name: RITTER2
Sterilizer ID: V2313228
02/24/2025 10:24 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 68.0 F 0.0
2:00 69.9 F 0.1
4:00 94.7 F 0.6
6:00 107.4 F 1.9
8:00 167.2 F 4.4
10:00 196.5 F 7.2
12:00 205.8 F 10.3
14:00 215.3 F 12.4
16:00 217.0 F 4.9
18:00 223.0 F 8.2
20:00 232.2 F 11.9
22:00 232.1 F 7.5
24:00 239.5 F 12.3
26:00 238.1 F 10.1
28:00 249.0 F 15.5
30:00 258.9 F 21.1
32:00 268.2 F 27.2

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.5
1:00 272.9 F 30.1
2:00 271.6 F 29.4
3:00 272.9 F 30.2
4:00 271.9 F 29.5

Min 271.5 F 29.4
Max 272.9 F 30.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.1 F 0.0
5:00 197.3 F 0.0
10:00 194.6 F 0.0
15:00 188.5 F 0.0
20:00 184.5 F 0.0
25:00 182.9 F 0.0
30:00 172.9 F 0.0

Drying Complete

FILLING: 0:55
HEATING: 33:23
STERILIZING: 4:00
VENTING: 3:50
DRYING: 30:00
TOTAL CYCLE: 01:12:07
CYCLE COMPLETE

02/24/2025 11:36 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____